



AURELIA® BOLD MAX

6mil, Nitrile Powder Free Examination Gloves

Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms. BS EN 455 1-4 Medical gloves for single use. EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 1186: 2002. Materials and articles in contact with foodstuffs.

AURELIA® GLOVES

A SUPERMAX HEALTHCARE BRAND

sales@supermax.ie

+353 15256724

Units 1001 – 1012, Building 1000,
Gateway Business Park,
New Mallow Road, Cork,
Ireland

aureliagloves.ie



10 x 50



UK
CA
0120

CE
0598



EN ISO 374-5: 2016
VIRUS



EN ISO 374-1: 2016/Type B
JKPT



M

L

XL

XXL

9789A7

9789A8

9789A9

9789A0

Sizes & Dimensions

Single wall thickness

Physical Properties



M 95+10 mm
L 110+10 mm
XL 120+10 mm
XXL > 120 mm



min. 240 mm
min. 240 mm
min. 240 mm
min. 240 mm



① min. 0.20 mm
② min. 0.17 mm
③ min. 0.08 mm

	Unaged	Aged
Tensile Strength	Min. 14.0 MPa	Min. 14.0 MPa
Elongation	Min. 500%	Min. 400%
Strength at Break	Min. 6N	Min. 6N



Nitrile



Powder Free



5 Years



Diamond



7.5g +/- 0.2g



Black

