



AURELIA® DELIGHT LITE PF

Powder Free Vinyl Examination Gloves



Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 455 1-4 Medical gloves for single use. BS EN 1186:2002 Materials and articles in contact with foodstuffs.

AURELIA® GLOVES

A SUPERMAX HEALTHCARE BRAND

sales@supermax.ie

+353 15256724

Units 1001 – 1012, Building 1000,
Gateway Business Park,
New Mallow Road, Cork
Ireland

aureliagloves.ie



S	M	L	XL
32226	32227	32228	32229

Sizes & Dimensions

Single wall thickness

Physical Properties

S 85+5 mm	min. 240 mm
M 95+5 mm	min. 240 mm
L 105+5 mm	min. 240 mm
XL > 115 mm	min. 240 mm

① min. 0.05 mm
② min. 0.05 mm
③ min. 0.04 mm

	Unaged	Aged
Tensile Strength	Min. 14.0 MPa	Min. 14.0 MPa
Elongation	Min. 400%	Min. 400%
Strength at Break	Min. 3.0N	Min. 3.0N



Vinyl



Powder Free



5 Years



Smooth



4.0g +/- 0.2g



Clear

