



AURELIA® AMAZING

2.2mil, Nitrile Powder Free Examination Gloves

Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms. BS EN 455 1-4 Medical gloves for single use. EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 1186: 2002. Materials and articles in contact with foodstuffs.

AURELIA® GLOVES

A SUPERMAX HEALTHCARE BRAND

sales@supermax.ie

+353 15256724

Units 1001 – 1012, Building 1000,
Gateway Business Park,
New Mallow Road, Cork,
Ireland

aureliagloves.ie



10 x 300

*



10 x 270



XS	S	M	L	*XL
92885	92886	92887	92888	92889

Sizes & Dimensions

Single wall thickness

Physical Properties

XS 70+10 mm	min. 240 mm
S 80+10 mm	min. 240 mm
M 95+10 mm	min. 240 mm
L 110+10 mm	min. 240 mm
XL > 110 mm	min. 240 mm



- ① min. 0.09 mm
- ② min. 0.05 mm
- ③ min. 0.05 mm

	Unaged	Aged
Tensile Strength	Min. 14.0 MPa	Min. 14.0 MPa
Elongation	Min. 500%	Min. 400%
Strength at Break	Min. 6N	Min. 6N



Nitrile



Powder Free



5 Years



Micro



3.0g +/- 0.2g



Purple Blue

