



# AURELIA® TRANSFORM

3.2mil, Nitrile Powder Free Examination Gloves

Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms. BS EN 455 1-4 Medical gloves for single use. EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 1186: 2002. Materials and articles in contact with foodstuffs.

## AURELIA® GLOVES

A SUPERMAX HEALTHCARE BRAND

sales@supermax.ie

+353 15256724

Units 1001 – 1012, Building 1000,  
Gateway Business Park,  
New Mallow Road, Cork,  
Ireland

[aureliagloves.ie](http://aureliagloves.ie)



10 x 200



XS	S	M	L	XL
98895	98896	98897	98898	98899

### Sizes & Dimensions

### Single wall thickness

### Physical Properties

XS 70+10 mm	min. 240 mm
S 80+10 mm	min. 240 mm
M 95+10 mm	min. 240 mm
L 110+10 mm	min. 240 mm
XL > 110 mm	min. 240 mm



- ① min. 0.10 mm
- ② min. 0.06 mm
- ③ min. 0.05 mm

	Unaged	Aged
Tensile Strength	Min. 14.0 MPa	Min. 14.0 MPa
Elongation	Min. 500%	Min. 400%
Strength at Break	Min. 6N	Min. 6N



Nitrile



Powder Free



5 Years



Finger Textured



3.6g +/- 0.2g



Ice Blue

